

Manual Supplement

Manual Title: VT650/VT900A Getting Started
Part Number: 5008662
Print Date: October 2021
Revision/Date: 2

Supplement Issue: 1
Date: 3/25
Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, TM-1007

The most recent version of the Getting Started Manual is Revision 2 and is available on the Fluke Biomedical website, this document supplements Revision 2 of the Getting Started Manual.

On page 5, under the section titled "Power On the Analyzer" add the following note to the end of the section:

Note:

A one-hour warm-up is recommended for maximum measurement accuracy and stability. Flow rate sensors are subject to offset drift during the warm-up period.